

## A quasioptical SIS receiver with normal metal tuning for the 800-900 GHz band

F. Schäfer\*, P. van der Wal\*, E. Kreysa\*, K.H. Gundlach\*\*

\* Max-Planck-Institut für Radioastronomie, D-53121 Bonn, Germany

\*\* Institut de Radioastronomie Millimetrique, F-38406 St. Martin d'Herès, France

We describe a quasi-optical SIS receiver designed for operation in the 800 - 900 GHz atmospheric window. First results obtained during an observation period in January 1997 at the Heinrich-Hertz-Telescope (HHT) on Mt. Graham, Arizona are presented.

The receiver employs the well established design of using a diffraction limited substrate-lens made of crystalline quartz fed by a double-slot antenna to couple to the incoming radiation [1]-[3]. In contrast to the design variant using silicon as lens material [4], our antenna configuration includes a moveable reflecting backplane to avoid losses due to the backward radiation into the vacuum side. Without backplane these losses would be higher as compared to silicon due to the lower refractive index of the quartz dielectric; on the other hand for the silicon lens an antireflection coating usually has to be applied to reduce the higher reflection loss of the silicon lens. Currently we use no antireflection coating on the quartz lens. A schematic drawing of our optical setup is shown in Fig. 1. It was designed using a theoretical method developed by Prof. V. Hansen and his group [5], which gives the antenna's impedance and its radiation characteristic into a dielectric half space. Results from the calculation of the double-slot input impedance versus reflector distance are shown in Fig. 2. The extended hyperhemispherical lens was shown to provide good quality amplitude and phase patterns by scaled measurements at microwave frequencies [6].

A photography illustrating the mechanical design of the mixer block is shown in Fig. 3. Only few mechanical components are necessary due to the use of a commercially available microwave substrate with thick copper groundplane [7]. It carries the lens assembly on the front side whereas the microwave and bias lines are etched on the substrate side which also supports the superconducting magnet with pole pieces and the reflecting backplane assembly.

The SIS junctions used for this experiment were produced in 1994 at IRAM-Grenoble (France) during a guest period in the junction laboratory. Because the desired operating band of the receiver is well above the gap frequency for niobium films, we decided to use a normal conducting embedding network for the Nb- (Al-oxide)-Nb junction [8]. DC measurements of sputtered aluminum films at 295 K and 4.2 K gave a resistivity near the bulk limit at room temperature and a resistivity ratio (RR) of roughly 11 for a variety of films. At high frequency and low temperature anomalous skin effect losses are the limiting factor for the achievable minimum RF loss in the Al films. According to the theory of metallic conduction, size effects impose an upper

limit on the DC conductivity of thin films. At low temperatures scattering of the conduction electrons from the film surface becomes important if the electron's bulk mean free path is comparable to the film thickness. The result is a maximum  $RR \approx 10$  for our film thickness of about 200nm, meaning that no further improvement of the RF surface resistance can be expected if this  $RR$  is reached. This is because in the pure bulk limit (clean films with thickness large compared to anomalous penetration depth) the RF surface resistance does not depend on the  $RR$  [9]. The mixer chip used for the measurements reported here is shown in Fig. 4. It employs the double-junction tuning circuit with antisymmetric feed configuration introduced by Zmuidzinas et al. [10]. We used the cross-line technique [11] to define junctions of roughly  $0.8 \times 0.8 \mu\text{m}^2$  by optical lithography on two tiny islands of trilayer deposited on the aluminum groundplane. The junction current density is about  $15 \text{ kA/cm}^2$ . A microphotograph of the junction area of the mixer chips is shown in Fig. 5. These photographs also show the two different thicknesses of the sputtered  $\text{SiO}_2$  dielectric layer that were used in the design. In order to realize line impedances of  $15\text{-}25 \Omega$  with acceptable linewidth (losses) for the antenna feed circuit,  $800 \text{ nm SiO}_2$  was used. In contrast, for the radial stub dc-blocks at the slots and for the tuning inductor between the junctions  $400 \text{ nm SiO}_2$  was used to realise a low impedance level. This thickness of the dielectric layer results naturally from the crossline-process because about  $200\text{nm}$  of  $\text{SiO}_2$  are needed for each of the two RIE and liftoff steps.

The receiver cryostat is a standard liquid helium bath type (IR-labs HD3-8). The measurements reported here were done at a mixer block temperature of  $4.2 \text{ K}$ . The cryostat's atmospheric window is made of  $100 \mu\text{m}$  mylar foil. IR filtering is done by means of a crystalline quartz plate with polyethylene antireflection coating on both sides at  $77 \text{ K}$  and a resonant fluorogold foil at  $4.2 \text{ K}$ . The cooled IF section consists of a 3 stage  $2 - 4 \text{ GHz}$  HEMT amplifier developed at MPIfR; no IF transformer between mixer output and amplifier was used.

During first laboratory tests of the receiver in 1995, a DSB receiver noise temperature of  $850 \text{ K}$  without any correction for losses had been obtained at a single frequency of  $803 \text{ GHz}$  using a stabilized FIR laser with  $^{15}\text{NH}_3$  isotope gas as local oscillator [12]. From subsequent measurements of the receiver bandpass shown in Fig. 6 using a Martin-Puplett interferometer [13], good heterodyne performance could be expected for the entire atmospheric window. When a solid state local oscillator system for the frequency range  $795 - 875 \text{ GHz}$  got available, we set up a tunable telescope system. The LO system consists of an InP Gunn-oscillator followed by a doubler and tripler stage [14]. A  $15 \mu\text{m}$  mylar beamsplitter couples the local oscillator power to the mixer, two ellipsoidal mirrors at ambient temperature are used to match the mixer beam to the telescope and to the dual-mode horn at the tripler output. Fig. 7 shows the DSB receiver noise temperatures achieved in the laboratory at  $4.2 \text{ K}$  bath temperature, no corrections for losses were made. The radiation temperatures of the loads at room and liquid nitrogen temperature were corrected according to the Callen and Welton formula [15].

The noise temperature of the IF chain including a cooled isolator was measured by means of a heated 50  $\Omega$  load to be 12.5 K in a measurement bandwidth of 60 MHz centered at 3 GHz. From these results a conversion loss of  $\approx 11$  dB was estimated at 803 GHz. The single measurement included in the figure indicates a DSB noise temperature of 680 K achieved at 810 GHz at the HHT. A plot of the corresponding conversion curves and the junction's IV-curve is given in Fig. 8. The improved sensitivity is attributed to the replacement of the cooled IF amplifier prior to the observation run and to the lower bath temperature of about 3.85 K on the mountain.

During our ten day technical observing run at the HHT in January, an astronomical verification of the receiver's performance could be obtained. As an example Fig. 9 shows the average of 4 continuum subscans across Mars (scanning in azimuth). Deconvolved from the intrinsic size of the planet, a FWHP main beam width of  $\Theta_{\text{mb}} \approx 9''$  was deduced from the measurements which corresponds to the telescope's diffraction limit. No evidence for side lobe or broad pedestal structure was found to a level of 10-13 dB from this and other measurements carried out. Heterodyne response is demonstrated in Fig. 10, showing the CO (7-6) transition detected towards the carbon rich circumstellar envelope IRC 10216. The integration time was 3 min only for this spectrum taken in rather poor atmospheric conditions ( $\tau_{\text{line-of-sight}} \geq 3$ ).

### **Acknowledgements**

We thank the staff of the workshop at MPIfR Bonn for careful machining of the mixerblock, and the staff of the junction laboratory at IRAM Grenoble for their support during our fabrication run at the junction laboratory. Part of this work was supported by Deutsche Agentur für Raumfahrt-Angelegenheiten (DARA) under contract FKZ 50 00 9102

### **References**

- [1] D.B. Rutledge, M. S. Muha, "Imaging Antenna Arrays", IEEE Trans. Antennas Prop., vol. 30, pp. 535-540, July 1982
- [2] M. Wengler, D. P. Woody, R.E. Miller, T.G. Phillips, "A Low Noise Receiver for Millimeter and Submillimeter Wavelengths", Int. J. IR and MM Waves, vol. 6, pp. 697-706, 1985
- [3] J. Zmuidzinas, H.G. LeDuc, "Quasi-Optical Slot Antenna SIS Mixers", IEEE Trans. MTT, vol. 40, pp1797-1804, Sept. 1992
- [4] J. Zmuidzinas, N.G. Ugras, D. Miller, M. Gaidis, H.G. LeDuc, J.A. Stern, "Low-Noise Slot Antenna SIS Mixers", IEEE Trans. Appl. Supercond., vol. 5, pp.3053ff., 1995

- [5] H. Aroudaki, T. Vaupel, V. Hansen, F. Schäfer, "Full-Wave Analysis of CPW Fed Slot Antennas for the Submillimeter Wave Region by the Spectral Domain Method", Prog. in Electromagnetic Research Symp. (PIERS), 1994, Nordwijk, Netherlands, Proc., pp. 1-4
- [6] F. Schäfer, E. Kreysa, T. Lehnert, K. H. Gundlach, "A Planar SIS Receiver with Logperiodic Antenna for Submillimeter Wavelengths", Proc. 4<sup>th</sup> Intl. Symp. on Space Terahertz Technologie, 1993, Los Angeles, CA
- [7] Duroid 6002, Rogers Corporation, Chandler, Arizona 85226, USA
- [8] R.L. Kautz, "Miniaturization of Normal-State and Superconducting Striplines", J. Res. Natl. Bur. Stand., vol. 84, pp 247-259, 1979
- [9] R.W. Keyes, E.P. Harris, K.L. Konnerth, "The Role of Low Temperatures in the Operation of Logic Circuitry", Proc. IEEE, vol. 58, pp. 1914-1932, 1970
- [10] J. Zmuidzinas, H. G. LeDuc, J. A. Stern, S. R. Cypher, "Two-Junction Tuning Circuits for Submillimeter SIS Mixers", IEEE Trans. MTT, vol. 42, pp. 698-706, April 1994
- [11] M. Voss, A. Karpov, K.H. Gundlach, "Submicron Nb-Aloxide-Nb Junctions for Frequency Mixers", Superconductor Science and Technol., vol. 6, pp.373ff, 1993
- [12] F. Schäfer, J. Kuhnert, E. Kreysa, K.H. Gundlach, "A quasioptical SIS receiver for the 800-900 GHz band", Proc. of the 3<sup>rd</sup> Intl. Workshop on Terahertz Electronics, Sept. 1995, Zermatt, Switzerland
- [13] J. Gromke, "Spektrale Kalibration von Radiometern im Submm Bereich mit Hilfe eines Martin-Puplett-Interferometers", diploma thesis, Universität Bonn, 1996
- [14] Radiometer Physics GmbH, Bergerwiesenstrasse 15, D-53340 Meckenheim, Germany
- [15] S.-K. Pan, A.R. Kerr, M.J. Feldman, A. Kleinsasser, J. Stasiak, R.L. Sandstrom, and W.J. Gallagher, "An 85-116 GHz SIS Receiver using Inductively Shunted Edge Junctions", IEEE Trans. MTT, vol. 37, pp. 580ff, 1989




Figure 1: Schematic diagram of the antenna configuration




Figure 2: Single slot input impedance for a double slot antenna




Figure 3: Photograph of the mixer block




Figure 4: Mixer chip for 800-900 GHz




Figure 5: Junction area of the mixer chip

FTS measurement of open structure SIS




Figure 6: FTS measurement of receiver bandpass




Figure 7: Receiver noise temperatures




Figure 8: IV and conversion curves for telescope result




Figure 9: Continuum subscans in azimuth across Mars




Figure 10: Spectrum of CO (7-6) towards IRC 10216