

18/1/62

Tested thermocouples.

They produce full scale deflection on 2" dia meter with 100 ma true couple as expected.


When proper 2" meter is replaced by a General Radio 3" meter the deflection is only 65% for same 100 ma true couple. Obviously the General Radio meter is not as sensitive as Vane 2" diameter meter. Use the proper meter.

10/1/62

The movement on Sangamo 4 ampere R.F. meter was disconnected from thermocouple. A movement from General Radio meter was then connected to thermocouple. This gave a full scale deflection of 3.25 amperes which is not significantly better than built in 4 ampere movement.

Detector

3/5/62



13/6/62

Calibration of Output meter using balanced input

Input 2.2 mVc

Input 0.8 V on half.

Volts

Input

across

Half

Output
ma△
ma

Freq.

Mc

right

side

left

side

				Output Ma
0.1	.093		1.	.972 .981
0.2	.131		1.6	.958 .988
0.3	.247		1.8	
0.4	.571		2.0	.946 .990
0.5	.500		2.2	.930 .992
0.6	.645		2.4	.919 .981
0.7	.785		2.6	.908 .976
0.8	.928		2.8	.903 .961
0.9	1.060		3.0	.888 .950

3 May 62

Calibration of 600- Ω detector at 50 cycles

Range	Low	Medium	High	High	3/3/64
Degeneration Resistor	170 ohms.	530 ohms	1380 ohms.	2070 ohms	
Ma. Output	RMS Volts Input.				
0	0	0	0	0	0
.1	.39	.52	1.72	.80	1.26
.2	.50	.74	2.81	1.23	2.10
.3	.60	.93	3.93	1.64	2.88
.4	.68	1.10	5.02	2.04	3.69
.5	.75	1.26	6.10	2.40	4.44
.6	.83	1.42	7.14	2.73	5.21
.7	.90	1.57	8.26	3.08	5.95
.8	.96	1.74	9.38	3.47	6.89
.9	1.02	1.89	10.40	3.80	7.65
1.0	1.09	2.05	11.47	4.19	8.47

High

7/10/64


4600- Ω degeneration
resistor

High

2/10/64

3150- Ω degeneration
resistor

Calibration of Voltmeter
3/5/62
4/2/64



1.0 Scale L }
2.0 Scale M }
3.0 Scale H }

Scale	Range (Microvolts)	Slope (Microvolts per Unit)	Y-intercept (Microvolts)
L	0 to 18	1.0	19.5
M	0 to 18	1.0	17.5
H	0 to 18	1.0	15.5

Calibration of 600 ohm Detector

Input 50 cycles. RMS

9 October 64


Range	Low	Middle	High
Degeneration Resistor ohms	170	963	3780

Microamperes
Output Input Volts

0	0	0	0
100	.38	.65	1.46
200	.49	.97	2.44
300	.57	1.75	3.39
400	.65	1.52	4.30
500	.73	1.77	5.23
600	.80	2.05	6.17
700	.86	2.30	7.10
800	.93	2.55	8.07
900	.99	2.79	9.03
1000	1.05	3.03	9.98

Calibration of 600 ohm Detector

9th October 1964



Output Microamperes