

18/6/63

7/7/63

Calibration of meter for reading vibrator voltage. Voltage Doubler circuit used

Volts AC RMS.	Volts DC Indicated	Volts DC Indicated
0	0	0
.1	.045	.100
.2	.100	.292
.3	.170	.465
.4	.248	.690
.5	.317	.890
.6	.400	1.103
.7	.475	1.300
.8	.552	1.520
.9	.645	1.725
1.0	.725	1.930
1.1	Bridge	2.150
1.2		2.355
1.3		2.60

Voltage Doubler

1.3
 2.6
 1.4

 104
 26

 364